

.

---

**From:** Nassif, Julianne (DPH)  
**Sent:** Monday, November 02, 2009 3:16 PM  
**To:** Piro, Peter (DPH); Salemi, Charles (DPH)  
**Subject:** RE: UPS Failure

I would keep using the instruments. I will check around to see if we have an extra UPS.  
The power is much more reliable than it has been in the past. Do you think we still need it?

---

**From:** Piro, Peter (DPH)  
**Sent:** Monday, November 02, 2009 11:13 AM  
**To:** Nassif, Julianne (DPH); Salemi, Charles (DPH)  
**Subject:** UPS Failure

Hello,

I have a UPS that appears to be having battery problems. Two GC/MS rough pumps are attached to this UPS. Do you want me to continue to run the instruments in bypass mode or should I take them down until I get the problem fixed? Charging without a load doesn't appear to be working so far.

-Peter